

Nederlands Meetinstituut

Member State  
The Netherlands

OIML Certificate N° R76/1992-NL-00.23  
Project number 10118378  
Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: W.A.C.M. van Leeuwen

### Applicant

Name: Mettler-Toledo GmbH  
Address: Im Langacher  
8606 Greifensee  
Switzerland

### Manufacturer of the certified pattern

Name: Mettler-Toledo GmbH  
Address: Im Langacher  
8606 Greifensee  
Switzerland

### Identification of the certified pattern

Type : AX..., MX..., UMX...

$n \leq 510\ 000$  divisions  
Max  $\leq 510$  g  
 $e = 1$  mg  
 $d \leq e \leq 10000d$

Temperature range  $+10\ ^\circ\text{C} / +30\ ^\circ\text{C}$

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the type-approval certificate and the description with number T5708 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organisation of Legal Metrology (OIML):

R76  
edition 1992

for accuracy class **I**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

Nederlands Meetinsituut  
Hugo de Grootplein 1  
3314 EG Dordrecht  
Telephone +31 78 6332332  
Telefax +31 78 6332309

NMi B.V. (Chamber of Commerce Haaglanden  
No. 27228701)


Subsidiary companies:  
NMi Certin B.V. (27233418)  
NMi Van Swinden Laboratorium B.V. (27228703)  
NMi International B.V. (27239176)

This document is issued under the provision that  
NMi. B.V. nor its subsidiary companies accept  
any liability.

Reproduction of the complete document is  
allowed. Parts of the document may only be  
reproduced after written permission



QUALIFIED  
BY STERLAB  
Reg. nr. L 029



Nederlands Meetinstituut

**Member State**  
The Netherlands

**OIML Certificate N° R76/1992-NL-00.23**  
Project number 10118378  
Page 2 of 2

The conformity was established by tests described in the associated test report:  
N° R76/1992-NL-00.23, that includes 41 pages.

The issuing authority  
W.A.C.M. van Leeuwen  
Manager Certification

31 October 2000

The OIML member  
G.J. Faber

31 October 2000

\*  
\*\*

---

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.